

DEPARTMENT OF THE INTERIOR
IRRIGATION OFFICE

RATING CURVE FOR GURLEYS LARGE ELECTRIC METER.

Pattern No 600

SHOWING

Revolutions per second and Velocity in feet per second

H.P. Davis
Chief Engineer.

Note - This Meter was in hard use for two seasons and shows a typical condition of the variance in rating curves. Note that with the old bearing the meter ran faster and that the introducing of a new bearing brought the rating curve closer to the standard.

No of rating	Date	Rated by	Legend	Remarks
0			a	Gurley Standard Curve
1	Feb 18 98	VAN B	b	Old bearing
2	Mar 20 98	VAN O	c	New bearing

