

TABLE VII.
TESTS ON HOLDERS USED TO TEST THE GAS METERS FOR REGULAR CITY SERVICES.
Holders were all Tested at about 1 in. Pressure.

No. 9 Holder in 1st Corner of Room.				Holder No. 14 in Centre of North Wall of Room.			
(1)	(2)	(3)	Error in Holder.	(1)	(2)	(3)	Error in Holder.
Part of Holder Scale Tested.	Volume shown by Holder Scale	Actual Volume of Air put into Holder.	Difference between Columns (2) and (3).	Part of Holder Scale Tested.	Volume shown by Holder Scale.	Actual volume of air put into Holder.	Difference between Columns (2) and (3).
	Cu. ft.	Cu. ft.	Cu. ft.		Cu. ft.	Cu. ft.	Cu. ft.
10.00-9.00	1.00	1.00	0	10.00-9.007	0.993	1.00	0.007
9.00-8.007	0.993	1.00	0.007	9.00-8.007	0.993	1.00	0.007
8.00-7.01	0.990	1.00	0.01	8.00-7.01	1.00	1.00	0
7.00-6.00	1.00	1.00	0	7.00-6.01	0.99	1.00	0.01
6.00-5.014	0.986	1.00	0.014	6.00-5.005	0.995	1.00	0.005
5.00-4.00	1.000	1.00	0	5.00-4.00	1.00	1.00	0
4.00-3.013	0.987	1.00	0.013	4.00-3.00	1.00	1.00	0
3.00-2.005	0.994	1.00	0.006	3.00-2.005	0.995	1.00	0.005
2.00-1.00	1.00	1.00	0	2.00-1.00	1.00	1.00	0
1.00-0	1.00	1.00	0	1.00-0	1.00	1.00	0
			Per cent.				Per cent.
			0				0.7
			0.7				0.7
			1.0				0
			0				1.0
			1.4				0.5
			0				0
			1.3				0
			0.6				0.5
			0				0
			0				0