

TABLE V.
RESULTS OF TESTS OF METERS SELECTED BY CITY'S REPRESENTATIVE.

Consumer	Date of Test.	Higher Pressure.			Lower Pressure.			Remarks.
		Pressure inches.	Cu. ft. air discharged per hr.	Percentage error of meter.	Pressure inches.	Cu. ft. air discharged per hr.	Percentage error of meter.	
Jas. Chan	June 15...	1.00	14	8.68 slow	Meter refused to work or pass air at Rejected. the half-inch pressure.			
M. Soloway	June 15...	1.00	15	No error...	.46	16	No error...	
E. G. Lynch	June 15...	1.00	14	3.29 slow	.48	14	4.90 slow...	Rejected.
J. A. Woodward	June 15...	1.00	14	1.13 slow	.48	14	3.52 slow...	
Geo. Yeo	June 15...	1.00	15	8.05 slow	.47	16	2.91 slow	Rejected.
J. D. McGill	June 15...	1.00	30	4.17 slow	.47	26	3.05 slow	Rejected.
J. W. Reeves	June 15...	1.00	32	2.72 slow	.48	25	3.10 slow...	
J. M. Percy	June 15...	1.00	32	0.86 fast...	.48	28	2.04 fast...	
Millburn Co.	June 15...	1.00	50	3.57 slow	.45	24	4.71 slow...	Rejected.
Reid Bros	June 16...	1.00	54	1.48 slow	.48	52	1.38 slow...	
S. Walker	June 16...	.98	55	0.40 slow	.48	51	1.11 fast	
F. R. Beetle	June 16...	.97	100	1.67 slow	.45	82	0.70 slow...	